

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	9.606 °C	10.1 °C	0.494 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.600 °C	26.1 °C	0.500 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.301 °C	39.8 °C	-0.501 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	18.648 %HR	20.3 %HR	1.652 %HR	± 3.0 %HR	± 1.9E+00 %HR
60 %HR	57.953 %HR	59.5 %HR	1.547 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.059 %HR	76.2 %HR	1.141 %HR	± 3.0 %HR	± 1.9E+00 %HR