

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	
10 °C	9.606 °C	9.9 °C	0.294 °C	±	0.5 °C
25 °C	25.600 °C	25.3 °C	-0.300 °C	±	0.5 °C
40 °C	40.301 °C	39.9 °C	-0.401 °C	±	0.5 °C
20 %HR	18.648 %HR	32.6 %HR	13.952 %HR	±	3.0 %HR
60 %HR	57.953 %HR	61.0 %HR	3.047 %HR	±	3.0 %HR
75 %HR	75.059 %HR	72.5 %HR	-2.559 %HR	±	3.0 %HR