

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	
10 °C	9.606 °C	9.8 °C	0.194 °C	±	0.5 °C
25 °C	25.600 °C	25.3 °C	-0.300 °C	±	0.5 °C
40 °C	40.301 °C	40.0 °C	-0.301 °C	±	0.5 °C
20 %HR	18.648 %HR	20.4 %HR	1.752 %HR	±	3.0 %HR
60 %HR	57.953 %HR	59.5 %HR	1.547 %HR	±	3.0 %HR
75 %HR	75.059 %HR	76.1 %HR	1.041 %HR	±	3.0 %HR