

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>
10 °C	9.606 °C	10.0 °C	0.394 °C	± 0.5 °C
25 °C	25.600 °C	25.8 °C	0.200 °C	± 0.5 °C
40 °C	40.301 °C	40.2 °C	-0.101 °C	± 0.5 °C
20 %HR	18.648 %HR	20.2 %HR	1.552 %HR	± 3.0 %HR
60 %HR	57.953 %HR	60.1 %HR	2.147 %HR	± 3.0 %HR
75 %HR	75.059 %HR	78.0 %HR	2.941 %HR	± 3.0 %HR