

Certificado No.: 66257

Equipo (Instrument): Wireless Datalogger

Fecha de Calibración: 2026-02-24

Marca (Brand): INDATA

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
22 °C	22.040 °C	22.4 °C	0.360 °C	± 1.0 °C	± 0.4 °C
25 °C	24.888 °C	25.5 °C	0.582 °C	± 1.0 °C	± 0.4 °C
31 °C	31.105 °C	31.0 °C	-0.060 °C	± 1.0 °C	± 0.4 °C
48 %HR	47.697 %HR	49.1 %HR	1.403 %HR	± 5.0 %HR	± 1.9 %HR
57 %HR	57.053 %HR	55.9 %HR	-1.153 %HR	± 5.0 %HR	± 1.9 %HR
67 %HR	66.603 %HR	65.6 %HR	-1.033 %HR	± 5.0 %HR	± 1.9 %HR