

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
0 °C	0.581 °C	1.0 °C	0.419 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.175 °C	20.4 °C	0.225 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.287 °C	35.6 °C	0.313 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	18.648 %HR	22.5 %HR	3.852 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.940 %HR	55.4 %HR	3.460 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	81.194 %HR	84.5 %HR	3.306 %HR	± 5.0 %HR	± 1.9E+00 %HR