

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
1.00 mm	1.00 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
1.50 mm	1.50 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
3.00 mm	3.00 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
5.00 mm	5.00 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
10.00 mm	10.00 mm	0 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
15.00 mm	15.01 mm	10 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
20.00 mm	20.01 mm	10 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
25.00 mm	25.01 mm	10 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
30.00 mm	30.02 mm	20 μm	$\pm 30 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$