

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
1.00 mm	1.01 mm	10 µm	± 30 µm	± 5.8E+00 µm
1.50 mm	1.51 mm	10 µm	± 30 µm	± 5.8E+00 µm
3.00 mm	3.01 mm	10 µm	± 30 µm	± 5.8E+00 µm
5.00 mm	5.01 mm	10 µm	± 30 µm	± 5.8E+00 µm
10.00 mm	10.01 mm	10 µm	± 30 µm	± 5.8E+00 µm
15.00 mm	15.01 mm	10 µm	± 30 µm	± 5.8E+00 µm
20.00 mm	20.01 mm	10 µm	± 30 µm	± 5.8E+00 µm
25.00 mm	25.02 mm	20 µm	± 30 µm	± 5.8E+00 µm
30.00 mm	30.02 mm	20 µm	± 30 µm	± 5.8E+00 µm