

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.173 °C	29.3 °C	-0.873 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.055 %HR	33 %HR	4.945 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	51.122 %HR	56 %HR	4.878 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	64.415 %HR	69 %HR	4.585 %HR	± 5.0 %HR	± 1.9E+00 %HR