

| Nominal (<i>Nominal</i>) | Patrón (<i>Standard</i>) | UBP (<i>UUT</i>) | Error (<i>Error</i>) | Exactitud (<i>Accuracy</i>) | Incertidumbre (<i>Uncertainty</i>) |
|-----------------------------------|-----------------------------------|---------------------------|-------------------------------|--------------------------------------|---|
| 30 °C | 30.173 °C | 29.2 °C | -0.973 °C | ± 1.0 °C | ± 4.0E-01 °C |
| 30 %HR | 28.055 %HR | 32 %HR | 3.945 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 51.122 %HR | 56 %HR | 4.878 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 70 %HR | 64.415 %HR | 69 %HR | 4.585 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |