

| Patrón (Standard) | UBP (UUT) | Error (Error) | E.M.P (Tolerance) | Incertidumbre (Uncertainty) | |
|--------------------------|---------------------|----------------------|--------------------------|------------------------------------|-----------------------|
| 52.65 μm | 52.86 μm | 0.21 μm | $\pm 1.6 \mu\text{m}$ | \pm | 5.8E-02 μm |
| 155.23 μm | 155.2 μm | -0.03 μm | $\pm 4.7 \mu\text{m}$ | \pm | 6.2E-02 μm |
| 257.12 μm | 257.2 μm | 0.08 μm | $\pm 7.7 \mu\text{m}$ | \pm | 7.0E-02 μm |
| 511.72 μm | 511.4 μm | -0.32 μm | $\pm 15.4 \mu\text{m}$ | \pm | 9.7E-02 μm |