

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
52.65 μm	52.98 μm	0.33 μm	$\pm 1.6 \mu\text{m}$	$\pm 5.8\text{E-}02 \mu\text{m}$
155.23 μm	154.8 μm	-0.43 μm	$\pm 4.7 \mu\text{m}$	$\pm 6.2\text{E-}02 \mu\text{m}$
257.12 μm	257.0 μm	-0.12 μm	$\pm 7.7 \mu\text{m}$	$\pm 7.0\text{E-}02 \mu\text{m}$
511.72 μm	511.8 μm	0.08 μm	$\pm 15.4 \mu\text{m}$	$\pm 9.7\text{E-}02 \mu\text{m}$