

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.051 °C	14.9 °C	-0.151 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.721 °C	30 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.618 °C	36 °C	-0.118 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	38.750 %HR	40 %HR	1.150 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	48.975 %HR	50.3 %HR	1.325 %HR	± 3.0 %HR	± 1.9E+00 %HR
90 %HR	89.875 %HR	91.3 %HR	1.425 %HR	± 3.0 %HR	± 1.9E+00 %HR