

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.005 mm	1.010 mm	5 µm	± 50.25 µm	±	5.8E-01 µm
1.5 mm	1.503 mm	3 µm	± 75. µm	±	5.8E-01 µm
2 mm	2.004 mm	4 µm	± 100 µm	±	5.8E-01 µm
4 mm	4.008 mm	8 µm	± 200 µm	±	5.9E-01 µm
5 mm	5.008 mm	8 µm	± 250 µm	±	5.8E-01 µm