

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
49.00 mmHg	50 mmHg	1.00 mmHg	± 5 mmHg	±	6.6E-01 mmHg
99.00 mmHg	100 mmHg	1.00 mmHg	± 5 mmHg	±	6.5E-01 mmHg
148.00 mmHg	150 mmHg	2.00 mmHg	± 5 mmHg	±	6.6E-01 mmHg
198.00 mmHg	200 mmHg	2.00 mmHg	± 5 mmHg	±	6.8E-01 mmHg
248.00 mmHg	250 mmHg	2.00 mmHg	± 5 mmHg	±	7.5E-01 mmHg
298.00 mmHg	300 mmHg	2.00 mmHg	± 5 mmHg	±	7.5E-01 mmHg