

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (M.P.E)</b>	<b>Incertidumbre (Uncertainty)</b>	
100.0209 ml	100 ml	-0.0209 ml	± 5.0 ml	±	3.5E-01 µl
250.0539 ml	250 ml	-0.0539 ml	± 5.0 ml	±	3.5E-01 µl
499.9273 ml	500 ml	0.0727 ml	± 5.0 ml	±	3.5E-01 µl
750.0694 ml	750 ml	-0.0694 ml	± 5.0 ml	±	3.5E-01 µl
1000.0040 ml	1000 ml	-0.0040 ml	± 5.0 ml	±	3.5E-01 µl