

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	
100.0310 ml	100 ml	-0.0310 ml	± 5.0 ml	±	3.5E-01 µl
250.0038 ml	250 ml	-0.0038 ml	± 5.0 ml	±	3.5E-01 µl
499.9073 ml	500 ml	0.0927 ml	± 5.0 ml	±	3.5E-01 µl
750.0564 ml	750 ml	-0.0564 ml	± 5.0 ml	±	3.5E-01 µl
999.9629 ml	1000 ml	0.0371 ml	± 5.0 ml	±	3.5E-01 µl