

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
0.00092 in	0.0009 in	-0.00002 in	± 0.0003 in	±	9.0E-05 in
0.00203 in	0.0020 in	-0.00003 in	± 0.0003 in	±	1.1E-04 in
0.00491 in	0.0049 in	-0.00001 in	± 0.0003 in	±	2.2E-04 in
0.00985 in	0.0098 in	-0.00005 in	± 0.0003 in	±	1.1E-04 in
0.01913 in	0.0191 in	-0.00003 in	± 0.0003 in	±	1.1E-04 in
0.03791 in	0.0379 in	-0.00001 in	± 0.0003 in	±	2.4E-04 in