

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
1.00 mm	1.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
1.25 mm	1.25 mm	0 µm	± 30 µm	±	5.8E+00 µm
1.50 mm	1.50 mm	0 µm	± 30 µm	±	5.8E+00 µm
2.00 mm	2.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
5.00 mm	5.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
10.00 mm	10.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
15.00 mm	15.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
20.00 mm	20.00 mm	0 µm	± 30 µm	±	5.8E+00 µm
25.00 mm	25.01 mm	10 µm	± 30 µm	±	5.8E+00 µm