

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
2 mm	1.5 mm	0 μm	$\pm 0.11 \mu\text{m}$	\pm	5.8E+01 μm
5 mm	5.0 mm	0 μm	$\pm 0.13 \mu\text{m}$	\pm	5.8E+01 μm
10 mm	10.0 mm	0 μm	$\pm 0.15 \mu\text{m}$	\pm	5.8E+01 μm
15 mm	14.8 mm	-200 μm	$\pm 0.18 \mu\text{m}$	\pm	5.8E+01 μm
20 mm	19.7 mm	-300 μm	$\pm 0.20 \mu\text{m}$	\pm	5.8E+01 μm
25 mm	24.7 mm	-300 μm	$\pm 0.23 \mu\text{m}$	\pm	5.8E+01 μm
50 mm	49.7 mm	-300 μm	$\pm 0.35 \mu\text{m}$	\pm	5.8E+01 μm
100 mm	99.7 mm	-300 μm	$\pm 0.60 \mu\text{m}$	\pm	5.8E+01 μm
200 mm	199.7 mm	-300 μm	$\pm 1.1 \mu\text{m}$	\pm	5.8E+01 μm