

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.452 °C	15.7 °C	0.248 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.420 °C	25.6 °C	0.180 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.253 °C	30.3 °C	0.047 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.599 %HR	64.3 %HR	4.701 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.691 %HR	68.2 %HR	3.509 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.895 %HR	72.2 %HR	2.305 %HR	± 3.0 %HR	± 1.9E+00 %HR