

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
15 °C	15.452 °C	15.8 °C	0.348 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.420 °C	25.8 °C	0.380 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.253 °C	30.7 °C	0.447 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.599 %HR	63.2 %HR	3.601 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.691 %HR	67.5 %HR	2.809 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.895 %HR	70.9 %HR	1.005 %HR	± 3.0 %HR	± 1.9E+00 %HR