

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	15.452 °C	15.8 °C	0.348 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.420 °C	25.7 °C	0.280 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.253 °C	30.4 °C	0.147 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	59.599 %HR	61.3 %HR	1.701 %HR	± 3.0 %HR	± 1.9E+00 %HR
65 %HR	64.691 %HR	65.6 %HR	0.909 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	69.895 %HR	73.4 %HR	3.505 %HR	± 3.0 %HR	± 1.9E+00 %HR