

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
# -30 °C	-30.224 °C	-30.0 °C	0.224 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.430 °C	-25.0 °C	0.430 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.232 °C	-15.0 °C	0.232 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.168 °C	-10.0 °C	0.168 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.845 °C	2.0 °C	0.155 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.921 °C	8.0 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C