

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
# -30 °C	-30.224 °C	-30.0 °C	0.224 °C	± 0.5 °C	± 4.0E-01 °C
-25 °C	-25.430 °C	-25.0 °C	0.430 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.232 °C	-15.0 °C	0.232 °C	± 0.5 °C	± 4.0E-01 °C
-10 °C	-10.168 °C	-10.0 °C	0.168 °C	± 0.5 °C	± 4.0E-01 °C
2 °C	1.845 °C	1.5 °C	-0.345 °C	± 0.5 °C	± 4.0E-01 °C
8 °C	7.921 °C	8.0 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C