

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
15 °C	14.770 °C	15.0 °C	0.230 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	24.758 °C	25.0 °C	0.242 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	29.832 °C	30.0 °C	0.168 °C	± 0.5 °C	± 4.0E-01 °C
60 %HR	57.124 %HR	55.0 %HR	-2.124 %HR	± 5.0 %HR	± 1.9E+00 %HR
65 %HR	63.086 %HR	60.5 %HR	-2.586 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	67.789 %HR	65.5 %HR	-2.289 %HR	± 5.0 %HR	± 1.9E+00 %HR