

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.1	0.10000	0.250	± 0.6	± 2.9E+00 mm					
0 a 0.2	0.20000	0.250	± 0.6	± 2.9E+00 mm					
0 a 0.3	0.30000	0.260	± 0.7	± 2.9E+00 mm					
0 a 0.4	0.40000	0.260	± 0.7	± 2.9E+00 mm					
0 a 0.5	0.50000	0.260	± 0.8	± 2.9E+00 mm					
0 a 0.6	0.60000	0.270	± 0.8	± 2.9E+00 mm					
0 a 0.7	0.70050	-0.230	± 0.8	± 2.9E+00 mm					
0 a 0.8	0.80050	-0.230	± 0.9	± 2.9E+00 mm					
0 a 0.9	0.90050	-0.220	± 0.9	± 2.9E+00 mm					
0 a 1	1.00100	-0.720	± 1	± 2.9E+00 mm					
