

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.75	0.750004	0.150	± 0.9	± 3.5E-01 mm					
0 a 0.8	0.800055	0.120	± 0.9	± 3.5E-01 mm					
0 a 0.9	0.900064	0.160	± 0.9	± 3.5E-01 mm					
0 a 1	1.000086	0.200	± 1	± 3.5E-01 mm					
0 a 1.2	1.200099	0.310	± 1	± 4.1E-01 mm					
0 a 1.4	1.400184	0.360	± 1.1	± 4.1E-01 mm					
0 a 1.6	1.600017	0.690	± 1.2	± 4.1E-01 mm					
0 a 1.85	1.850111	0.830	± 1.3	± 4.1E-01 mm					
0 a 2	2.000079	1.020	± 1.4	± 4.1E-01 mm					
