

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
15 °C	15.021 °C	14.9 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.028 °C	25.1 °C	0.072 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.268 °C	32.3 °C	0.032 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.314 %HR	20.3 %HR	-0.014 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	47.4 %HR	-3.283 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.936 %HR	67.2 %HR	-2.736 %HR	± 5.0 %HR	± 1.9E+00 %HR