

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.330 °C	-24.9 °C	0.430 °C	± 0.5 °C	± 4.0E-01 °C
-15 °C	-15.152 °C	-14.8 °C	0.352 °C	± 0.5 °C	± 4.0E-01 °C
0 °C	-0.211 °C	0.2 °C	0.411 °C	± 0.5 °C	± 4.0E-01 °C