

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
15 °C	15.021 °C	15.1 °C	0.079 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.028 °C	25.0 °C	-0.028 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.268 °C	32.3 °C	0.032 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.314 %HR	22.7 %HR	2.386 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	48.3 %HR	-2.383 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.936 %HR	72.5 %HR	2.564 %HR	± 5.0 %HR	± 1.9E+00 %HR