

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
15 °C	15.021 °C	15.2 °C	0.179 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.028 °C	25.1 °C	0.072 °C	± 0.5 °C	± 4.0E-01 °C
32 °C	32.268 °C	32.4 °C	0.132 °C	± 0.5 °C	± 4.0E-01 °C
20 %HR	20.314 %HR	22.2 %HR	1.886 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.683 %HR	49.0 %HR	-1.683 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.936 %HR	72.5 %HR	2.564 %HR	± 5.0 %HR	± 1.9E+00 %HR