

Patrón (Standard)	UBP (UT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
180.075 s	180.067 s	-0.008 s	-0.004%	± 0.01%	± 5.0E-02 s
600.124 s	600.12 s	-0.004 s	-0.001%	± 0.01%	± 5.0E-02 s
1800.451 s	1800.45 s	-0.001 s	-0.0001%	± 0.01%	± 5.1E-02 s
3600.124 s	3600.12 s	-0.004 s	-0.0001%	± 0.01%	± 1.7E+00 s