

<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )	
180.075 s	180.067 s	-0.008 s	-0.004% ±	0.01% ±	1.8E-01 s
600.124 s	600.12 s	-0.004 s	-0.001% ±	0.01% ±	1.8E-01 s
1800.451 s	1800.45 s	-0.001 s	-0.0001% ±	0.01% ±	1.8E-01 s
3600.124 s	3600.12 s	-0.004 s	-0.0001% ±	0.01% ±	1.8E-01 s