

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.213 °C	11.3 °C	-0.913 °C	± 1.0 °C	± 4.0E-01 °C
27 °C	27.415 °C	27.9 °C	0.485 °C	± 1.0 °C	± 4.0E-01 °C
42 °C	42.396 °C	42.1 °C	-0.296 °C	± 1.0 °C	± 4.0E-01 °C
52 %HR	52.400 %HR	53.0 %HR	0.600 %HR	± 5.0 %HR	± 1.9E+00 %HR
72 %HR	72.493 %HR	69.0 %HR	-3.493 %HR	± 5.0 %HR	± 1.9E+00 %HR
92 %HR	91.920 %HR	93.0 %HR	1.080 %HR	± 5.0 %HR	± 1.9E+00 %HR