

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
12 °C	12.213 °C	11.8 °C	-0.413 °C	± 1.0 °C	± 4.0E-01 °C
27 °C	27.415 °C	27.3 °C	-0.115 °C	± 1.0 °C	± 4.0E-01 °C
42 °C	42.396 °C	41.6 °C	-0.796 °C	± 1.0 °C	± 4.0E-01 °C
52 %HR	52.400 %HR	52.0 %HR	-0.400 %HR	± 5.0 %HR	± 1.9E+00 %HR
72 %HR	72.493 %HR	73.0 %HR	0.507 %HR	± 5.0 %HR	± 1.9E+00 %HR
92 %HR	91.920 %HR	93.0 %HR	1.080 %HR	± 5.0 %HR	± 1.9E+00 %HR