

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
2 °C	1.985 °C	1.5 °C	-0.485 °C	± 1.0 °C	± 4.0E-01 °C
6 °C	5.996 °C	5.6 °C	-0.396 °C	± 1.0 °C	± 4.0E-01 °C
8 °C	7.811 °C	7.6 °C	-0.211 °C	± 1.0 °C	± 4.0E-01 °C
55 %HR	55.202 %HR	59.0 %HR	3.798 %HR	± 5.0 %HR	± 1.9E+00 %HR