

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
30 °C	30.253 °C	30.6 °C	0.347 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.719 %HR	32.0 %HR	3.281 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.578 %HR	54.3 %HR	4.722 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.895 %HR	70.0 %HR	0.105 %HR	± 5.0 %HR	± 1.9E+00 %HR