

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.253 °C	29.8 °C	-0.453 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.719 %HR	31 %HR	2.281 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.578 %HR	51 %HR	1.422 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.895 %HR	70 %HR	0.105 %HR	± 5.0 %HR	± 1.9E+00 %HR