

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.253 °C	29.8 °C	-0.453 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	28.719 %HR	27 %HR	-1.719 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.578 %HR	47 %HR	-2.578 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.895 %HR	68 %HR	-1.895 %HR	± 5.0 %HR	± 1.9E+00 %HR