

<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )	
3600.221 s	3600 s	-0.221 s	-0.006%	± 0.01 %	± 1.0E+03 s
3600.219 s	3600 s	-0.219 s	-0.006%	± 0.01 %	± 1.0E+03 s
3600.234 s	3600 s	-0.234 s	-0.006%	± 0.01 %	± 1.0E+03 s