

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.163 °C	30.3 °C	0.137 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.366 %HR	31.8 %HR	1.434 %HR	± 3 %HR	± 1.9E+00 %HR
50 %HR	50.335 %HR	52.2 %HR	1.865 %HR	± 3 %HR	± 1.9E+00 %HR
70 %HR	70.652 %HR	72.5 %HR	1.848 %HR	± 3 %HR	± 1.9E+00 %HR