

---

<b>Patrón (Standard)</b>		<b>UBP (UT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
100 Hz	1.000 g-pk	0.910 g-pk	-0.090 g-pk	-	± 0.021 g-pk
100 Hz	2.000 g-pk	1.950 g-pk	-0.050 g-pk	-	± 0.021 g-pk
100 Hz	3.000 g-pk	2.960 g-pk	-0.040 g-pk	-	± 0.021 g-pk
100 Hz	4.000 g-pk	3.930 g-pk	-0.070 g-pk	-	± 0.021 g-pk
100 Hz	5.000 g-pk	4.960 g-pk	-0.040 g-pk	-	± 0.021 g-pk