

Patrón (Standard)		UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
50 Hz	1.000 g-pk	0.91 g-pk	-0.09 g-pk	-	± 0.021 g-pk
100 Hz	1.000 g-pk	0.95 g-pk	-0.05 g-pk	-	± 0.021 g-pk
160 Hz	1.000 g-pk	0.93 g-pk	-0.07 g-pk	-	± 0.021 g-pk
300 Hz	1.000 g-pk	0.96 g-pk	-0.04 g-pk	-	± 0.021 g-pk
500 Hz	1.000 g-pk	0.95 g-pk	-0.05 g-pk	-	± 0.021 g-pk
100 Hz	# 2.000 g-pk	1.92 g-pk	-0.08 g-pk	-	± 0.021 g-pk
100 Hz	# 3.000 g-pk	2.90 g-pk	-0.10 g-pk	-	± 0.021 g-pk
100 Hz	# 4.000 g-pk	3.93 g-pk	-0.07 g-pk	-	± 0.021 g-pk
100 Hz	# 5.000 g-pk	4.96 g-pk	-0.04 g-pk	-	± 0.021 g-pk