

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
10 °C	10.308 °C	10.5 °C	0.192 °C	± 0.5 °C	± 4.0E-01 °C
20 °C	20.135 °C	20.5 °C	0.365 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.463 °C	30.5 °C	0.037 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.090 °C	40.5 °C	0.410 °C	± 0.5 °C	± 4.0E-01 °C