

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
10 °C	10.308 °C	10.5 °C	0.192 °C	± 0.5 °C	± 2.4E-01 °C
20 °C	20.135 °C	20.5 °C	0.365 °C	± 0.5 °C	± 1.7E-01 °C
30 °C	30.463 °C	30.5 °C	0.037 °C	± 0.5 °C	± 1.6E-01 °C
40 °C	40.090 °C	40.5 °C	0.410 °C	± 0.5 °C	± 1.6E-01 °C