

| <b>Nominal (Nominal)</b> | <b>Patrón (Standard)</b> | <b>UBP (UUT)</b> | <b>Error (Error)</b> | <b>Exactitud (Accuracy)</b> | <b>Incertidumbre (Uncertainty)</b> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 12 °C                    | 12.213 °C                | 12.0 °C          | -0.213 °C            | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 27 °C                    | 27.415 °C                | 27.0 °C          | -0.415 °C            | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 42 °C                    | 42.396 °C                | 42.0 °C          | -0.396 °C            | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 52 %HR                   | 52.400 %HR               | 54.5 %HR         | 2.100 %HR            | ± 3.0 %HR                   | ± 1.9E+00 %HR                      |
| 72 %HR                   | 72.493 %HR               | 74.0 %HR         | 1.507 %HR            | ± 3.0 %HR                   | ± 1.9E+00 %HR                      |
| 92 %HR                   | 91.920 %HR               | 93.0 %HR         | 1.080 %HR            | ± 3.0 %HR                   | ± 1.9E+00 %HR                      |