

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
12 °C	12.213 °C	12.0 °C	-0.213 °C	± 0.5 °C	± 4.0E-01 °C
27 °C	27.415 °C	27.0 °C	-0.415 °C	± 0.5 °C	± 4.0E-01 °C
42 °C	42.396 °C	42.0 °C	-0.396 °C	± 0.5 °C	± 4.0E-01 °C
52 %HR	52.400 %HR	54.5 %HR	2.100 %HR	± 3.0 %HR	± 1.9E+00 %HR
72 %HR	72.493 %HR	72.5 %HR	0.007 %HR	± 3.0 %HR	± 1.9E+00 %HR
92 %HR	91.920 %HR	94.0 %HR	2.080 %HR	± 3.0 %HR	± 1.9E+00 %HR