

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
35 °C	35.357 °C	35.7 °C	0.343 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.202 %HR	28.8 %HR	-1.402 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	50.8 %HR	0.772 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.568 %HR	70.0 %HR	0.432 %HR	± 5.0 %HR	± 1.9E+00 %HR