

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
35 °C	35.357 °C	36.0 °C	0.643 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.202 %HR	30.0 %HR	-0.202 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.028 %HR	51.0 %HR	0.972 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	69.568 %HR	70.8 %HR	1.232 %HR	± 5.0 %HR	± 1.9E+00 %HR